•	04-15-	65	<b>4</b>
O I PE Applica	ition No.	Applicant(s)	. /
10/646	,002	RIEBAU ET AL.	
Office Action Summary APR 1 4 2005 Examir	er	Art Unit	
Chuck		3676	
The MAILING DATE of this communication of Period for Reply	the cover sheet with the d	correspondence ad	ddress (
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET THE MAILING DATE OF THIS COMMUNICATION.  - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no after SIX (6) MONTHS from the mailing date of this communication.  - If the period for reply specified above is less than thirty (30) days, a reply within the set of NO period for reply is specified above, the maximum statutory period will apply and Failure to reply within the set or extended period for reply will, by statute, cause the any reply received by the Office later than three months after the mailing date of this earned patent term adjustment. See 37 CFR 1.704(b).	event, however, may a reply be tir tatutory minimum of thirty (30) day will expire SIX (6) MONTHS from polication to become ABANDONE	nely filed  s will be considered time the mailing date of this of D (35 U.S.C. § 133).	ly. communication.
Status			
1) Responsive to communication(s) filed on  2a) This action is <b>FINAL</b> . 2b) This action is 3) Since this application is in condition for allowance exceded closed in accordance with the practice under <i>Ex parte</i> 0.	ot for formal matters, pro		e merits is
Disposition of Claims			
<ul> <li>4)  Claim(s) 1 is/are pending in the application.</li> <li>4a) Of the above claim(s) is/are withdrawn from constant of the state of</li></ul>			
Application Papers			
9) The specification is objected to by the Examiner.  10) The drawing(s) filed on is/are: a) accepted or accepted or Applicant may not request that any objection to the drawing(s).  Replacement drawing sheet(s) including the correction is requested.  11) The oath or declaration is objected to by the Examiner.	be held in abeyance. See tired if the drawing(s) is obj	e 37 CFR 1.85(a). ected to. See 37 Cl	• •
Priority under 35 U.S.C. § 119			
12) Acknowledgment is made of a claim for foreign priority of a) All b) Some * c) None of:  1. Certified copies of the priority documents have be 2. Certified copies of the priority documents have be 3. Copies of the certified copies of the priority documents have be application from the International Bureau (PCT Ref. * See the attached detailed Office action for a list of the certified copies.	een received. een received in Application nents have been receive ule 17.2(a)).	on No ed in this National	Stage
Attachment(s)			
1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal Pa 6) Other:	te	D-152)

# Notice of References Cited Notice of References Cited PAPPLICATION Application/Control No. 10/646,002 Examiner Chuck Mah PATENT DOCUMENTS Applicant(s)/Patent Under Reexamination RIEBAU ET AL. Page 1 of 1

*		Document Number Country Code-Number-Kind Code	Date MM-YYYY	Name	Classification
	Α	US-6,546,594	04-2003	Wills, Steven	16/422
	В	US-6,512,158	01-2003	Dobos, John A.	. 602/41
	С	US-6,491,996	12-2002	Digangi, Mary Ann	428/43
	D	US-6,385,817	05-2002	Johnson, Ron D.	16/431
	Ε	US-6,187,696	02-2001	Lim et al.	442/77
	F	US-5,722,672	03-1998	Frederick, Sonya	280/33.992
	G	US-5,686,706	11-1997	Wurzburger, Isaac	181/131
	Н	US-5,480,377	01-1996	Cartmell et al.	602/48
	ı	US-5,364,677	11-1994	Sendziak, Walter	428/40.1
	J	US-4,981,737	01-1991	Rico, Nicholas	428/41.7
	К	US-3,935,890	02-1976	Dolnick, Seymour S.	4/661
	L	US-			
	М	US-			

## FOREIGN PATENT DOCUMENTS

*		Document Number Country Code-Number-Kind Code	Date MM-YYYY	Country	Name	Classification
	N	None				
	0					
	Р					
	α					
	R					
	S				·	
	Т					

## **NON-PATENT DOCUMENTS**

*		Include as applicable: Author, Title Date, Publisher, Edition or Volume, Pertinent Pages)
	C	
	٧	
	w	
	х	

<sup>\*</sup>A copy of this reference is not being furnished with this Office action. (See MPEP § 707.05(a).) Dates in MM-YYYY format are publication dates. Classifications may be US or foreign.

Application/Control Number: 10/646,002

d±Unit: 3676

#### **DETAILED ACTION**

# Claim Rejections - 35 USC § 112

U.S.C. 112, second paragraph.

The claim(s) are narrative in form and replete with indefinite and functional or operational language. The structure which goes to make up the device must be clearly and positively specified. The structure must be organized and correlated in such a manner as to present a complete operative device. The claim(s) must be in one sentence form only. Note the format of the claims in the patent(s) cited.

Further, it is not clear, from both the claim and the disclosure, whether "non-porous polyethylene plastic foam pad, non-latex rubber material" is defined as two separate laminated structures (laminated polyethylene plastic foam or laminated rubber) or a one-piece laminate having both foam and rubber.

## Claim Rejections - 35 USC § 103

- 2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
  - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- 3. Claim 1 is rejected under 35 U.S.C. 103(a) as being unpatentable over Lim et al. (6,187,696) in view of Dolnick (3,935,890).
  - '696 discloses a protective cover as claimed but for using non-porous polyethylene plastic foam and non-latex rubber material for forming the laminate.

Art Unit: 3676

It would have been obvious to one having ordinary skill in the art at the time the invention was made to form the laminate of '696 with non-porous polyethylene plastic foam and non-latex rubber material, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin, 125 USPQ 416.* 

Further, '696 does not show an adhesive surface for attachment to an equipment. '890 teaches a protective cover having an adhesive surface for attaching to a surface. It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the protective cover of '696 with an adhesive surface so that the protective cover can be easily be attached to any surface as desired.

## Conclusion

4. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Chuck Mah whose telephone number is (703) 308-0676. The examiner can normally be reached on 5/4-9.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Heather Shackelford can be reached on (703) 308-2978. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Application/Control Number: 10/646,002

Art Unit: 3676

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Chuck Matr

Primary Examiner Art Unit 3676

CM

# I claim:

1. The invention relates to a protective covering that will provide a sanitary cover to aid in the prevention of the spread of viruses and bacteria among persons sharing medical, physical therapy and physical equipment, devices and accessories, allowing everyday usage of medical, physical therapy and physical exercise equipment by multiple users. The I C WRAP is composed of a 1/16 inch thick Polyethylene foam pad, backed with clear HDPE adhesive film. Producing a non-slip, equipment-contacting side, and a sanitary surface on the patient contact surface. The protective covers for use with or without being provided in a roll of covers that have perforations at predetermined spaced intervals that allow the user to tear off one or more sections to achieve a cover of a desired size or in flat sheets.